SNAS SLOVENSKÁ NÁRODNÁ AKREDITAČNÁ SLUŽBA

CERTIFICATE
OF ACCREDITATION

No. K-151
dated 30.06.2026

The Slovak National Accreditation Service issues a Certificate of Accreditation to an accredited body pursuant to Section 26 par.6 of Act No. 53/2023 Coll. on Accreditation of Conformity Assessment Bodies (hereinafter referred to as the "Accreditation Act").

VYPRO, s.r.o.
K zábraniu 1653, 911 01 Trenčín
IČO: 36 705 942

Organizational unit performing the activity of the Accredited Body:
Calibration laboratory VYPRO

Workplace of the Accredited Body:
K zábraniu 1653, 911 01 Trenčín

Identification number of the Accredited Body: 758/K-151

Area of accreditation: Calibration laboratory

The accredited body demonstrated its competence to perform the accredited activity fulfilling the accreditation requirements of **ISO/IEC 17025: 2017** Standard when performing calibration of measuring instruments of force and force sensors within the scope of accreditation stated in the Annex of this Certificate of Accreditation. The Annex shall form an integral part of the Certificate of Accreditation.

Number and date of issue of the accreditation decision: No. 758/12723/2026/1 dated 15.06.2026

Validity of the accreditation decision:
The accreditation decision No. 758/12723/2026/1 dated 15.06.2026 takes effect on 30.06.2026 and is valid until 30.06.2031.

SLOVENSKÁ NÁRODNÁ AKREDITAČNÁ SLUŽBA 2 SNAS

Štefan Král
director

SNAS is signatory to the EA MLA and ILAC MRA.

record number: 12723/706845

Annex to the Certificate of Accreditation No. K-151 dated 30.06.2026.

The Annex is an integral part of the Certificate of Accreditation

Scope of Accreditation

Accredited body: VYPRO, s.r.o.
K zábraniu 1653, 911 01 Trenčín

Organizational unit performing the activity of the accredited body:
Calibration laboratory VYPRO

Place of performance of the accredited body:
K zábraniu 1653, 911 01 Trenčín

Identification number of the accredited body: 758/K-151

Laboratory with fixed scope

Item	Kind of measuring instrument/ measurement means	Measured quantity	Measurement range	Expanded uncertainty U (k=2)	Established Methods		Other specifications
					Kind/ Principle	Identification	
1	Measuring instruments of force and force sensors	force	(0.1 - 2) kN	0.1 % RD [1] tension 0.1 % RD [1] compression	Direct comparison with standard	PP - Force gauges EN ISO 376	In laboratory
		force	(2 - 200) kN	0.07 % RD [1] tension 0.07 % RD [1] compression			

Notes:
[1] RD – Reading value
PP – Working procedure

***

SLOVENSKÁ NÁRODNÁ AKREDITAČNÁ SLUŽBA
SNAS

record number: 12723/706845

SNAS SLOVENSKÁ NÁRODNÁ AKREDITAČNÁ SLUŽBA

OSVEDČENIE
O AKREDITÁCII

č. K-151
zo dňa 30.06.2026

Slovenská národná akreditačná služba vydáva podľa § 26 ods. 6 zákona č. 53/2023 Z. z. o akreditácii orgánov posudzovania zhody (ďalej len „zákon o akreditácii") osvedčenie o akreditácii akreditovanej osobe

VYPRO, s.r.o.
K zábraniu 1653, 911 01 Trenčín
IČO: 36 705 942

Organizačná zložka vykonávajúca činnosť akreditovanej osoby:
Kalibračné laboratórium VYPRO

Miesto výkonu činnosti akreditovanej osoby:
K zábraniu 1653, 911 01 Trenčín

Identifikačné číslo akreditovanej osoby: 758/K-151

Oblasť akreditácie: Kalibračné laboratórium

Akreditovaná osoba preukázala spôsobilosť vykonávať akreditovanú činnosť plnením akreditačných požiadaviek normy **ISO/IEC 17025: 2017** na výkon kalibrácie silomerov a snímačov sily podľa rozsahu akreditácie uvedeného v prílohe tohto osvedčenia o akreditácii. Príloha tvorí neoddeliteľnú súčasť osvedčenia o akreditácii.

Číslo a dátum vydania rozhodnutia o akreditácii: č. 758/12723/2026/1 zo dňa 15.06.2026

Čas platnosti rozhodnutia o akreditácii:
Rozhodnutie o akreditácii č. 758/12723/2026/1 zo dňa 15.06.2026 nadobúda účinnosť dňa 30.06.2026 a je platné do dňa 30.06.2031.

SLOVENSKÁ NÁRODNÁ AKREDITAČNÁ SLUŽBA 2 SNAS

Ing. Štefan Král, PhD.
riaditeľ

SNAS je signatárom EA MLA a ILAC MRA.

číslo reg. záznamu: 12723/706845